

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3D00.01
-----------------------------------------------------------------------------------	-------------------------------------------------------	-----------

## TECHNOVIEW® Apixaban CONT High 5x1 mL

**REF** 5090270-RUO

**LOT** 3D02C00.02

 2022-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		204.4 – 340.6 ng/mL	272.5	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	204.4 – 340.6 ng/mL	268.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Lucas HOLOVICZ

29. APR. 2021

Datum / Date:

**Freigabe von**  
**Approved by**

  
Suzana GRESAK PODAKAR

29. APR. 2021

Datum / Date: