
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00.01
---	---	-----------

Coagulation Reference, 5x1mL

REF 5220110-RUO
LOT 1R03C00.05
 2022-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron [®] alpha	Fibrinogen, Clauss method, diluted plasma CCN LOT 1P02C01	2.57 – 3.85 g/L	3.79	OK
	Fibrinogen, Clauss method, diluted plasma CCA LOT 3P91B01	0.93 – 1.39 g/L	1.27	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Suzana GRESAK PODAKAR

29. APR. 2021

Datum / Date:

Freigabe von
Approved by

Michael KAFKA

29. APR. 2021

Datum / Date: