

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE:	2023-07
PRODUCT NUMBER:	82098539		
LOT NUMBER:	N0119866		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	74	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	65.7	72.7	nkat/vial	66.4	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 1-27-21
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 1/29/21
 PRINT SIGNATURE

TITLE: QC Technical Supervisor