





# Certificate of Analysis

## Ecarin Diluent (20 mL)


Ready to use solution: 0.05 mol/L Tris pH 7.6, with NaCl, CaCl<sub>2</sub>, bovine serum albumin and polyethylene glycol.

 <b>REF</b>	Catalog number	ED0413-20
 <b>LOT</b>	Batch code	EDR1751
	Use by	2023 06
	Storage temperature	2-8°C

Analysis	Specification	Result
pH (neat solution)	7.60 ± 0.06 at 25°C (Temp. coefficient = -0.028 per °C)	7.65 at 25°C
Conductivity (neat solution)	16.5 ± 1.6 mS/cm at 25°C (Temp. coefficient = +2.1% per °C)	16.1 mS/cm at 25°C

The above Lot# has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications.

All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each kit package.

  
Pia Bryngelhed  
Quality Manager, Rossix AB

February 24, 2020  
Date



Rossix AB  
Taljegårdsgatan 3B  
SE-431 53 Molndal  
Sweden

Phone: +46 31 706 8965  
Fax: +46 31 706 8966  
[info@rossix.com](mailto:info@rossix.com)  
[www.rossix.com](http://www.rossix.com)