
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ1P00.01
---	---	-----------


### Coagulation Control N 5x1 mL


REF      5020040-RUO  
LOT      1P02C01.04  
      2022-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	FII clotting	78.8 – 131.3 %	98.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


**Durchführung von**  
**Prepared by**

  
Lucas HODOVICZ

  
Datum / Date:

**Freigabe von**  
**Approved by**

  
Christoph ENNE

  
Datum / Date: