

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3U00.01
---	---	-----------

TECHNOVIEW Arixtra Control L 6x1 mL

REF 5090012-RUO

LOT 3U02C00.03

 2022-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Richtwert Expected value Ceveron® alpha	Anti Xa Arixtra	0.37 – 0.69 µg/mL	0.53	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa Arixtra	0.37 - 0.69 µg/mL	0.56	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lucas HOLOVICZ

12. AUG. 2021
Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

12. AUG. 2021
Datum / Date: