
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0500.01
---	---	-----------

**Technochrom AT III manual method Kit ~39 T.**

REF            5340225-RUO

LOT             0511C00.02

            2022-10-31

Inhalt Content	Reagenz Reagent	LOT	
1x	Substrate Th-1 10µmol QC	8B12C00	2024-05-31
1x	Technochrom ATIII Reagent A1 20µmol	7U11C00	2023-07-31
1x	Technochrom ATIII Puffer A1 100mL	9871B00	2022-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	CCN LOT 1P02C01	75.8 – 126.3%	94.8	OK
	CCA LOT 3P91B01	33.8 – 56.3%	39.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Lucas HOLOVICZ

26. JULI 2021

Datum / Date:

**Freigabe von**  
**Approved by**

  
Michael KAFKA

26. JULI 2021

Datum / Date: