



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRB0.04  
Att.09  
RB16C00.02

## TECHNOZYM vWF:CBA Collagen Typ III ELISA Kit 96T.

REF	5450301-RUO
LOT	RB16C00.02
	2022-09-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND02C0	2022-09-30
1x	0.5 mL	vWF Calibrator 2	2ND02C0	2022-09-30
1x	0.5 mL	vWF Calibrator 3	3ND02C0	2022-10-31
1x	0.5 mL	vWF Calibrator 4	4ND02C0	2022-10-31
1x	0.5 mL	vWF Calibrator 5	5ND02C0	2022-09-30
1x	0.5 mL	vWF High Control	AND02C0	2022-10-31
1x	0.5 mL	vWF Low Control	BND02C0	2022-10-31
1x	12 mL	TMB Substrate	RKN11C0	2023-02-28
1x	12 mL	Stop solution	RKK12C0	2023-05-31
1x	80 mL	Wash buffer concentrate	RNA12C0	2023-06-30
1x	90 mL	Incubation buffer	RNB12C0	2023-06-30
1x	-	vWF:CBA Plate Collagen Typ III	RNC13C0	2023-07-31
1x	0.3 mL	CBA Conjugate	RNE11C0	2023-03-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Ca11-Ca15 ≥ 0.7	1.5		OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND02C0 0.90 - 1.50 IU/mL	1.41 IU/mL		OK
	vWF Low Control LOT BND02C0 0.21 - 0.49 IU/mL	0.33 IU/mL		OK



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRB0.04

Att.09

RB16C00.02

Durchgeföhrt von  
Prepared by

Michael KAFKA

05. AUG. 2021

Datum / Date:

Freigabe von  
Approved by

Christoph ENNE

06. AUG. 2021

Datum / Date: