

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE:	2023-07
PRODUCT NUMBER:	82265063		
LOT	N0714395		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Analysis with COATEST APC R:					PASS/FAIL
APT Time <u>CaCl₂</u>	28	34	seconds	31	PASS
<u>APC/CaCl₂</u>	90	--	seconds	123	PASS
APC Ratio	3.5	4.4	--	4.0	PASS

Analysis with COATEST APC R V:					
APT Time <u>CaCl₂</u>	--	38	seconds	32	PASS
<u>APC/CaCl₂</u>	75	--	seconds	98	PASS
APC Ratio	2.7	3.2	--	3.1	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel
PRINT

N. M. Patel

SIGNATURE

DATE: 7/30/2021

7-30-21

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza
PRINT

Donna Tencza

SIGNATURE

DATE: 7/30/2021

7/30/21

TITLE: Senior QC Analyst