

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2023-06
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N0614239		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0094	PASS
Intercept	--	0.250	--	0.170	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range					
Slope	0.020	0.060	--	0.043	PASS
Intercept	--	0.406	--	0.263	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS:

DESCRIPTION	LOT NUMBER	VOLUME
COATEST SP FVIII CACL2	N0613858	6 mL
COATEST SP FVIII FACTOR IXa + FACTOR X	N0614238	3 mL
COATEST SP FVIII S-2765 + I-2581	N0613857	12 mL
COATEST SP FVIII BUFFER	N0613875	20 mL
COATEST SP FVIII PHOSPHOLIPID	N0613873	2 mL

PREPARED BY: Nehal Patel *N.m. Patel* DATE: 7/26/2021 *7-26-21*
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 7/26/2021 *7/26/2021*
 PRINT SIGNATURE

TITLE: Technical Supervisor