

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	S-2302, 25 mg Substrate	EXPIRATION DATE:	2024-06
PRODUCT NUMBER:	82034039		
LOT NUMBER:	N0210537		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.4	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

PREPARED BY: Bhumi Vasoya *Bhumi Vasoya* DATE: 02-24-2021  
 PRINT SIGNATURE

TITLE: QC Analyst

APPROVAL SIGNATURE: Debra Harz *Debra Harz* DATE: 3-31-2021  
 PRINT SIGNATURE

TITLE: QC Tech Supervisor