

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7D00AZ
---	---	----------

## TECHNOTHROMBIN® TGA RC Low 5x0.5mL

REF      5006212  
LOT      7D11C00.02  
      2023-03-31

<b>Inhalt</b> <b>Content</b>	<b>Reagenz</b> <b>Reagent</b>	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x0.5 mL	TGA RC Low	7D11C00	2023-03-31

<b>Analytische Prozedur</b> <b>Analytical Procedure</b>		<b>Anforderung</b> <b>Requirement</b>	<b>Ergebnis</b> <b>Result</b>	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z11C00	175 - 291	194.5	OK
	TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2811C00	139 - 274	142.1	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI 

19. AUG. 2021  
Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA 

20. AUG. 2021  
Datum / Date: