

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3D00.01
---	---	-----------


**TECHNOVIEW® Apixaban CONT High 5x1 mL**

REF      5090270-RUO  
LOT      3D02C00.03  
      2022-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		204.4 – 340.6 ng/mL	272.5	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	204.4 – 340.6 ng/mL	268.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


**Durchführung von  
Prepared by**

  
Lucas HOLOVICZ

  
Datum / Date:

**Freigabe von  
Approved by**

  
Alessandro BENIGNI

  
Datum / Date: