

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0W00.01
---	---	-----------


Technochrom AT III Kit ~ 100 T.

REF 5340224-RUO
LOT 0W14C00.02
 2023-11-30

Inhalt Content	Reagenz Reagent	LOT	
2x25 mL	NaCl Solution 0,9 %	9502C00	2025-11-30
1x	ATIII Substrate Th-1 10 µmol	8B04C00	2023-11-30
1x	Technochrom ATIII Reagent A2 20 µmol	7S11C00	2024-02-29

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	CCN LOT 1P02C01	75.8 – 126.3 %	104.3	OK
Ceveron® alpha Chromogenic Substrate F IIa	CCA LOT 3P91B01	33.8 – 56.3 %	46.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by



Lavdri M. NASUFI

05. MAI 2021
Datum / Date:

Freigabe von
Approved by



Suzana GRESAK PODAKAR

06. MAI 2021
Datum / Date: