



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0J00.02

Att. 14

0J13C00.02

## TECHNOCHROM C1-INH Reagent Kit

REF	5345003-RUO
LOT	0J13C00.02
	2023-05-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	1mL	Coagulation Control N for C1-INH	1P12CC1H	2023-05-31
1x	1mL	Coagulation Control A for C1-INH	3P11CC1H	2023-05-31
1x	see label	Coagulation Reference for C1-INH	1R12CC1H	2023-05-31
1x	3mL	Substrate f. C1 18 µmol	8C01C00	2025-06-30
1x	3mL	C1-Esterase	7B11C00	2023-05-31
1x	25mL	Probenpuffer A	9M11C00	2026-05-31
1x	20mL	Reaktionspuffer B	9N12C00	2026-08-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Content	according to packaging SOP
	Vial label	confirmed
	Vial barcode	OK
	Box label	confirmed
	Box barcode	OK
	Package insert	confirmed
Kontrollwiederfindung am Ceveron <sup>®</sup> alpha	<b>Control N</b> LOT 1P12CC1H	74.0   -   123.3   % 96.6   %   OK
Recovery of controls on Ceveron <sup>®</sup> alpha	<b>Control A</b> LOT 3P11CC1H	32.5   -   54.1   % 42.9   %   OK

Durchgeführt von  
Prepared by

Christoph ENNE

Datum / Date:

Freigabe von  
Approved by

Suzana GRESAK PODAKAR

Datum / Date: