



Certificate of Analysis

READS
Monoclonal Free
Protein S Test Kit

Kit Lot #: MFPS-204
Plate Lot #: FSW-071
Product #: 051-001
Expiration Date: 2022-08-22

Quality Control Results

Testing Date: 2021-04-14
Manufacture Date: 2021-02-23

Absorbance	Value	Specification	Result by Assay
Reference Plasma 100% OD	98 % of Normal	≥ 1.0	1.104
Reagent Blank		≤ 0.250	0.083
Reference Curve R ²		≥ 0.95	0.99353

Control Response	Label Range	Result by Assay
Normal Control	81 – 148 %	130.5 %
Deficient Control	14 – 46 %	36.8 %

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

QA Approval By: Matt Evans

Signed:  2021-04-20

Title: Senior Quality Associate

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Phone: (303) 457-4345
Fax: (303) 457-4519