

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	<u>Antithrombin 10 x 10 IU</u>	EXPIRATION DATE:	<u>2023-09</u>
PRODUCT NUMBER:	<u>82072039</u>		
LOT NUMBER:	<u>N0311306</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
ANTITHROMBIN, 10 IU (654023)					
ANTITHROMBIN ACTIVITY	9	11	IU/vial	10	PASS

All human source material used to manufacture this product have been tested according to FDA guidelines and found to be non-reactive for: HBsAg, antibodies to HIV1 and 2, and antibodies to HCV

PREPARED BY: Bhumi Vasoya Bhumi Vasoya DATE: 03-29-2021
 PRINT SIGNATURE

TITLE: QC Analyst

APPROVAL SIGNATURE: Debra Harz Debra Harz DATE: 3-29-21
 PRINT SIGNATURE

TITLE: QC Tech supervisor