

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE:	2023-11
PRODUCT NUMBER:	82098539		
LOT NUMBER:	N0513173		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
<b>FACTOR Xa, 71 nkat (650515)</b>					
ACTIVITY	61	81	nkat/vial	65	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	65.7	72.7	nkat/vial	68	PASS

**KIT COMPONENTS :**

NAME	LOT NUMBER	VOLUME
FXa, 71 nkat	N0513173	2 ml

PREPARED BY: Bhumi Vasoya  
PRINT Bhumi Vasoya SIGNATURE

DATE: 06/15/2021

TITLE: QC Analyst - II

APPROVAL SIGNATURE: Debra Harz  
PRINT Debra Harz SIGNATURE

DATE: 6/15/21

TITLE: QC Tech Supervisor