

# CERTIFICATE OF ANALYSIS



PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2023-03
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0311713	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
<b>HIGH RANGE</b>					
Slope	0.33	0.57	--	0.43	PASS
Intercept	---	0.12	--	0.06	PASS
r	0.990	---	--	0.998	PASS
<b>LOW RANGE</b>					
Slope	2.8	6.9	--	4.8	PASS
Intercept	---	0.21	--	0.11	PASS
r	0.990	---	--	1.000	PASS

**KIT COMPONENTS**

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0311707	3 mL
COAMATIC FVIII S-2765 + I-2581	N0311711	6 mL
COAMATIC FVIII BUFFER	N0311447	24 mL

· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Nehal Patel *N.M. Patel*      DATE: 5/10/21  
 PRINT      SIGNATURE      5/10/2021

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*      DATE: 5/10/21  
 PRINT      SIGNATURE      5/10/2021

TITLE: Technical Supervisor