

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2024-04
PRODUCT NUMBER:	82032439		
LOT NUMBER:	N0412236		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.2	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 5-19-21
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Harz *Debra Harz* DATE: 5/25/21
 PRINT SIGNATURE

TITLE: QC Technical Supervisor