

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2024-04
PRODUCT NUMBER:	82141339		
LOT:	N0412615		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.3	PASS

PREPARED BY: Bhumi Vasoya *Bhumi Vasoya* DATE: 6/8/2021 *06/08/2021*
 PRINT SIGNATURE

TITLE: QC Analyst - II

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 6/8/2021 *6/8/21*
 PRINT SIGNATURE

TITLE: Senior QC Analyst