



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRXXAZ

TECHNOZYM® t-PA Combi Actibind ELISA Kit 96T

REF TC16000
LOT 160003C1.01
🕒 2023-03-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x	anti t-PA Plate	RGN11C0	2023-04-30
1x0.3mL	POX anti t-PA	RKM11C0	2023-03-31
1x80mL	Wash buffer Concentrate	RNA13C0	2023-09-30
1x90mL	Incubation buffer	RNB12C0	2023-06-30
1x12mL	TMB Substrat	RKN12C0	2023-07-31
1x12mL	Stop solution	RKK12C0	2023-05-31
1x0.5mL	Standard t-PA	RAW11C0	2023-08-31
1x20mL	Detection mix buffer- TRIS	RDA11C0	2023-09-30
1x	Detection mixture	RBA12C0	2023-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1 activity	> 0.5	0.52	OK
	Calibrator 5 activity	< 0.1	0.02	OK
	Calibrator 1 antigen	> 1.0	1.84	OK
	Calibrator 5 antigen	< 0.1	0,00	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

15. Okt. 2021

Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI *Alessandro Benigni*

15-Okt-2021

Datum / Date: