
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7D00AZ
---	---	----------

## TECHNOTHROMBIN® TGA RC Low 5x0.5mL

REF    5006212  
LOT    7D11C00.03  
    2023-03-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x0.5 mL	TGA RC Low	7D11C00	2023-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z11C00	175 - 291	194.5	OK
	TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2811C00	139 - 274	142.1	OK

**Durchführung von**  
**Prepared by**

Michael KAFKA



02. NOV. 2021

Datum / Date:

**Freigabe von**  
**Approved by**

Fikret NASUFI



02. NOV. 2021

Datum / Date: