

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ1Z00AZ</p>
---	---	---

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320

LOT 1Z14C00.01

⌚ 2022-10-31

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1mL	TGA Control High	1Z14C00	2022-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC Low (nM Peak Thrombin) LOT 7D11C00	74 - 240	107.4	OK
	TGA RC High (nM Peak Thrombin) LOT 7F11C00	281 - 468	341.4	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

25. OKT. 2021

Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

25. OKT. 2021

Datum / Date: