
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0700AZ
---	---	----------

Ceveron® TGA CAL Set

REF 5006347-RUO
LOT 0712C00.02
 2023-03-31

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	Ceveron TGA CAL BUF	9S11C00	2023-03-31
1x0.5mL	Ceveron TGA CAL	5Y13C	2023-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kalibrierung / Calibration Gegen / against WHO 01/580		report	884	OK	
Verwendbarkeit / usability Gerät / Analyzer: Ceveron 100 series	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	1.0	OK	
Kontrollwiederfindung Recovery of controls LOT 1Y13C00	Peak Thrombin nM	Ceveron® TGA RB LOT 7T11C	83 – 147	128	OK
Kontrollwiederfindung Recovery of controls LOT 2712C00	Peak Thrombin nM	Ceveron® TGA RB LOT 7T11C	56 – 100	92	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Michael KAFKA




22. NOV. 2021

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI



22. NOV. 2021

Datum / Date: