

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ1P00.01
---	---	-----------

### Coagulation Control N 5x1 mL

REF      5020040-RUO

LOT      1P02C01.07

      2022-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	F IX clotting	87.0 – 145.0 %	120.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Christoph ENNE

05. NOV. 2021

Datum / Date:

**Freigabe von**  
**Approved by**

  
Smilja DURANOVIC

09. NOV. 2021

Datum / Date: