

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRB0.04
		Att.09
		RB19C00.02

TECHNOZYM vWF:CBA Collagen Typ III ELISA Kit 96T.

REF	5450301-RUO
LOT	RB19C00.02
	2023-07-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND02C1	2023-11-30
1x	0.5 mL	vWF Calibrator 2	2ND02C1	2023-11-30
1x	0.5 mL	vWF Calibrator 3	3ND02C1	2023-11-30
1x	0.5 mL	vWF Calibrator 4	4ND02C1	2023-11-30
1x	0.5 mL	vWF Calibrator 5	5ND02C1	2023-11-30
1x	0.5 mL	vWF High Control	AND02C1	2023-11-30
1x	0.5 mL	vWF Low Control	BND02C1	2023-11-30
1x	12 mL	TMB Substrate	RKN12C0	2023-07-31
1x	12 mL	Stop solution	RKK13C0	2023-10-31
1x	80 mL	Wash buffer concentrate	RNA13C0	2023-09-30
1x	90 mL	Incubation buffer	RNB13C0	2023-09-30
1x	-	vWF:CBA Plate Collagen Typ III	RNC14C0	2023-09-30
1x	0.3 mL	CBA Conjugate	RNE12C0	2023-09-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Ca1-CaI5	≥ 0.7	1.5	OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND02C1	0.90 - 1.50 IU/mL	1.27 IU/mL	OK
	vWF Low Control LOT BND02C1	0.21 - 0.49 IU/mL	0.36 IU/mL	OK



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRB0.04

Att.09

RB19C00.02

Durchgeföhrt von
Prepared by

Fikret NASUFI *Fikret Nasufi*

25. NOV. 2021

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

25. NOV. 2021

Datum / Date: