



TECHNOCHROM C1-INH Reagent Kit

REF 5345003-RUO
 LOT 0J14C00.02
 2023-05-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	1mL	Coagulation Control N for C1-INH	1P12CC1H	2023-05-31
1x	1mL	Coagulation Control A for C1-INH	3P11CC1H	2023-05-31
1x	see label	Coagulation Reference for C1-INH	1R12CC1H	2023-05-31
1x	3mL	Substrate f. C1 18 µmol	8C01C00	2025-06-30
1x	3mL	C1-Esterase	7B12C00	2023-09-30
1x	25mL	Probenpuffer A	9M11C00	2026-05-31
1x	20mL	Reaktionspuffer B	9N12C00	2026-08-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Kontrollwiederfindung am Ceveron [®] alpha Recovery of controls on Ceveron [®] alpha	Control N LOT 1P12CC1H	74.0 - 123.3 %	93.4 %	OK
	Control A LOT 3P11CC1H	32.5 - 54.1 %	52.2 %	OK

Durchgeführt von
Prepared by


Lucie HOLONACZ

11. NOV 2021
Datum / Date:

Freigabe von
Approved by


Christoph ENNE

11. NOV 2021
Datum / Date: