## **CERTIFICATE OF ANALYSIS**



Instrumentation Laboratory

A Werfen Company

PRODUCT NAME : PRODUCT NUMBER :

COAMATIC FACTOR VIII

82258563

LOT NUMBER:

N0916390

EXPIRATION DATE: 2023-09

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	PASS/FAIL
		MIN	MAX	UNITS		
HIGH RANGE						
	Slope	0.33	0.57		0.37	PASS
	Intercept		0.12		0.11	PASS
	r	0.990			0.996	PASS
_OW RANGE						
	Slope	2.8	6.9		5.1	PASS
	Intercept		0.21		0.13	PASS
	r	0.990			0.999	PASS

 NAME
 LOT NUMBER
 VOLUME

 COAMATIC FVIII FACTOR REAGENT
 N0916376
 3 mL

 COAMATIC FVIII S-2765 + I-2581
 N0916194
 6 mL

 COAMATIC FVIII BUFFER
 N0916060
 24 mL

have been manufactured a	nd tested according to the requ	irements of their Manufacturing and Quality Control	procedures and have been f	ound to meet all
PREPARED BY:	Girlie Reinoso	k. Keinoso	DATE:	10/8/2021
	PRINT	SIGNATURE		and the second s
TITLE:	QC Analyst III			10/8/21
APPROVAL SIGNATURE:	D. Teneza	P. Genera	DATE:	10/8/2021
	PRINT	SIGNATURE		
TITLE:	Senior Group Leader			