CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation Laboratory
A Werfen Company

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2024-09
PRODUCT NUMBER:	82032439		
LOT NUMBER:	N0916380		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Girlie Reinoso (PRINT) G. Reinoso (SIGNATURE) DATE: 12/22/2021

TITLE: QC Analyst III

APPROVAL SIGNATURE: Donna Tencza (PRINT) D. Tencza (SIGNATURE) DATE: 12/22/21

TITLE: QC Senior Analyst