

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2023-12
PRODUCT NUMBER:	82266863	
LOT NUMBER:	N1218391	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Analysis with COATEST APC R V:					
APT Time <u>CaCl₂</u>	27	40	seconds	32	PASS
APC Ratio	1.4	1.9	--	1.8	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Bhumi Vasoya *Bhumi Vasoya*
 PRINT SIGNATURE

DATE: 1/27/2022 *01/27/2022*

TITLE: QC Analyst II

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza*
 PRINT SIGNATURE

DATE: 1/27/2022 *1/27/22*

TITLE: Senior Group Leader