

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

Instrumentation  
Laboratory  
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2023-10
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N1016952	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.41	PASS
Intercept	---	0.12	--	0.10	PASS
r	0.990	---	--	0.997	PASS
LOW RANGE					
Slope	2.8	6.9	--	5.2	PASS
Intercept	---	0.21	--	0.13	PASS
r	0.990	---	--	0.999	PASS

**KIT COMPONENTS**

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N1016742	3 mL
COAMATIC FVIII S-2765 + I-2581	N1016951	6 mL
COAMATIC FVIII BUFFER	N1016950	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Nehal Patel *N-M-Patel* DATE: 11/17/2021 *11-17-21*  
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 11/17/2021 *11/17/21*  
 PRINT SIGNATURE

TITLE: Senior Group Leader