

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2023-11
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N1117871		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0103	PASS
Intercept	--	0.250	--	0.186	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range					
Slope	0.020	0.060	--	0.044	PASS
Intercept	--	0.406	--	0.275	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS:

DESCRIPTION	LOT NUMBER	VOLUME
COATEST SP FVIII CACL2	N1117539	6 mL
COATEST SP FVIII FACTOR IXa + FACTOR X	N1117702	3 mL
COATEST SP FVIII S-2765 + I-2581	N1117870	12 mL
COATEST SP FVIII BUFFER	N1117853	20 mL
COATEST SP FVIII PHOSPHOLIPID	N1117600	2 mL

PREPARED BY: Nehal Patel *N-m-Patel*
 PRINT SIGNATURE

DATE: 12/7/2021 *12-7-21*

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza*
 PRINT SIGNATURE

DATE: 12/7/2021 *12/7/21*

TITLE: Senior Group Leader