

# CERTIFICATE OF ANALYSIS



DiaPharma Group, Inc.  
8948 Beckett Road, West Chester, OH 45069

<b>Product:</b>	DiaPharma Factor X kit
<b>Item Number:</b>	DPGFX
<b>Vendors:</b>	R <sup>2</sup> Diagnostics, Chromogenix
<b>Lot #:</b>	<b>DPGFX055</b>
<b>Exp. Date:</b>	<b>2023-09-30</b>

Kit Components:			
Name	Lot Number	Exp. Date	Volume (1 vial each)
FXa Chromogenic Substrate	N0815163	2024-08-31	25 mg
Russell's Viper Venom	DR-1141	2023-09-30	<1 mg
CaCl <sub>2</sub>	DC-1141	2023-09-30	20 ml
Tris Buffer	DB-1141	2023-09-30	100 ml

Test Description	Specifications		Results		Disposition
	Min	Max	Units	Value	
Slope	1.0e-3	4.0e-3	NA	2.894E-03	Pass
Intercept	-.09	.09	NA	1.827E-02	Pass
Regression Coefficient	R >	0.990	NA	0.998	Pass

Prepared By: Ryan Yeager Date: 2022-03-14  
Approved By: [Signature] Date: 03/18/2022