

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0300.01
---	---	-----------


## Technoview® Rivaroxaban CAL Set High 5x1 mL

REF      5090171-RUO  
LOT        0321C00.02  
        2024-02-29

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x1 mL	Technoview Rivaroxaban CAL1 High	5121C00	2024-02-29
1x1 mL	Technoview Rivaroxaban CAL2 High	5221C00	2024-02-29
1x1 mL	Technoview Rivaroxaban CAL3 High	5321C00	2024-02-29
1x1 mL	Technoview Rivaroxaban CAL4 High	5421C00	2024-02-29
1x1 mL	Technoview Rivaroxaban CAL5 High	5521C00	2024-02-29

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control Medium <span style="border: 1px solid black; padding: 2px;">LOT</span> 3511C00	78.1 – 130.1 ng/mL	89.5	OK
	Rivaroxaban Control High <span style="border: 1px solid black; padding: 2px;">LOT</span> 3311C00	241.8 – 403.0 ng/mL	315.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Lucas HOLOVICZ

12. APR. 2022

Datum / Date:

**Freigabe von**  
**Approved by**

  
Suzana GRESAK PODAKAR

12. APR. 2022

Datum / Date: