



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ4C00.02

Protein C Mangelplasma 2 x 1 mL

REF 5346806-RUO

LOT 4C01C0E.01

 2024-04-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.8	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P11C00	69.0 – 115.0 %	79.3 OK
	CCA LOT 3P22C00	8.3 – 13.8 %	10.4 OK
Etiketten Überprüfung Label Verification		OK	

Durchführung von
Prepared by


Lavdrin NASUFI

27. APR. 2022

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

28. APR. 2022

Datum / Date: