

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	<u>Antithrombin 1 x 25 IU</u>	EXPIRATION DATE:	<u>2023-12</u>
PRODUCT NUMBER:	<u>81079639</u>		
LOT NUMBER:	<u>N1218393</u>		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Antithrombin Activity	25	30	IU/vial	30	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0029	PASS
Intercept @ 37°C	0.40	0.60		0.43	PASS
r ²	0.990	--		0.999	PASS

PREPARED BY: Girlie Reinoso
PRINT

G. Reinoso
SIGNATURE

DATE: 3/21/2022 3/21/2022

TITLE: QC Analyst III

APPROVAL SIGNATURE: Donna Tencza
PRINT

Donna Tencza
SIGNATURE

DATE: 3/21/2022 3/21/2022

TITLE: Senior Group Leader