

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE:	2024-01
PRODUCT NUMBER:	82098539		
LOT NUMBER:	N0714741		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	73	PASS
STANDARD CURVE (r ²)	0.990	1.000	-	1.000	PASS
CONTROL RECOVERY	65.7	72.7	nkat/vial	67.5	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	VOLUME
FXa, 71 nkat	N0714741	2 ml

PREPARED BY: Maria Somoza
PRINT

Maria Somoza
SIGNATURE

DATE: 8/10/2021

TITLE: QC Analyst II

APPROVAL SIGNATURE: Debra Harz
PRINT

Debra Harz
SIGNATURE

DATE: 8/10/2021

TITLE: QC Tech Supervisor