

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3300.01
-----------------------------------------------------------------------------------	-------------------------------------------------------------	-----------

## TECHNOVIEW® Rivaroxaban Control High 5x1 mL

**REF**      5090174-RUO

**LOT**      3311C00.05

      2023-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		241.8 – 403.0 ng/mL	322.4	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW® anti-Xa	241.8 – 403.0 ng/mL	301.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Lucas HOLOVICZ

05. MAI 2022  
Datum / Date:

**Freigabe von**  
**Approved by**

  
Christoph ENNE

05. MAI 2022  
Datum / Date: