

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE:	2024-02
PRODUCT NUMBER:	82265063		
LOT	N0229758		

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>	--	38	seconds	30	PASS
	<u>APC/CaCl₂</u>	75	--	seconds	95	PASS
APC Ratio		2.7	3.2	--	3.1	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Bhumi Vasoya Bhumi Vasoya
PRINT SIGNATURE

DATE: 2/15/2022 02-15-2022

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.M. Patel
PRINT SIGNATURE

DATE: 2/15/2022 2-15-22

TITLE: Senior QC Analyst

QCM00C00423L-1, Rev. 02