

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2025-02
PRODUCT NUMBER:	82032439		
LOT NUMBER:	N0229905		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

PREPARED BY: María Somoza *María Somoza* DATE: 3/8/2022 *3/8/2022*  
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 3/8/2022 *3-8-22*  
 PRINT SIGNATURE

TITLE: Senior QC Analyst