

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3300.01
---	---	-----------

**TECHNOVIEW® Rivaroxaban Control High 5x1 mL**

**REF** 5090174-RUO

**LOT** 3311C00.05

 2023-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		241.8 – 403.0 ng/mL	322.4	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW® anti-Xa	241.8 – 403.0 ng/mL	301.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
Lucas HOLOVICZ

05. MAI 2022

Datum / Date:

**Freigabe von  
Approved by**

  
Christoph ENNE

05. MAI 2022

Datum / Date: