



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Ca-Chloridlösung 25 mmol/L 25 mL

REF 5277017-RUO

LOT 9C03CB0.02

2025-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
CaCl ₂ Gehalt CaCl ₂ Content		23.5 – 26.5 mmol/L	25.2	OK	
Verwendbarkeit, Usability Ceveron® alpha	Siron LS	TCN LOT 1A05C00	22.5 – 30.5 sec	26.5	OK
		TCA LOT 3A91B00	57.0 – 77.2 sec	67.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Alessandro BENIGNI

13-Jun-2022
Datum / Date:

Freigabe von
Approved by

Lucas HOLOWICZ

14-Jun-2022
Datum / Date: