



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQxUT0.03

Att.12

7UT12C0.02

TECHNOZYM® ADAMTS13 Activity CON Set 2x0.5 mL

REF 5450763-RUO
LOT 7UT12C0.02
⌚ 2023-07-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	⌚
1x	0.5mL	ADAMTS13 Activity High Control	RUY12C0	2023-07-31
1x	0.5mL	ADAMTS13 Activity Low Control	RUZ12C0	2023-07-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Kontrollwiederfindung Recovery of controls	ADAMTS13 Activity High Control LOT RUY12C0	0.56 - 0.94 IU/mL	0.77 IU/mL	OK
	ADAMTS13 Activity Low Control LOT RUZ12C0	0.14 - 0.32 IU/mL	0.24 IU/mL	OK

Durchgeführt von
Prepared by

Fikret NASUFI *Fikret Nasufi*

06. SEP. 2021

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *M. Kafka*

06. SEP. 2021

Datum / Date: