



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

9613C00.02

Ceveron Cleaning Solution 25 mL

REF 9820300-RUO
LOT 9613C00.02
⌚ 2022-12-31

| Analytische Verfahren Analytical Procedures | | Anforderung Requirement | Ergebnis Result | |
|--|---------------------------------|----------------------------|--------------------|----|
| visuelle Kontrolle visual inspection | Cap color | white | confirmed | OK |
| | Vial type | white plastic vial | confirmed | OK |
| | Vial label | according to packaging SOP | confirmed | OK |
| | Vial barcode | according to packaging SOP | confirmed | OK |
| | Box label | according to packaging SOP | confirmed | OK |
| | Box barcode | according to packaging SOP | confirmed | OK |
| | Appearance | clear liquid | confirmed | OK |
| Kontrollwiederfindung Fibrinogen am Ceveron alpha | Control N LOT 1P02C01 | 2.57 - 3.85 g/L | 3.77 g/L | OK |
| Recovery of controls Fibrinogen on Ceveron alpha | Control A LOT 3P02C00 | 1.14 - 1.72 g/L | 1.58 g/L | OK |

Durchgeführt von
Prepared by

Smilja DURANOVIC
Smilja Duranovic

20. JAN. 2022

Datum / Date:

Freigabe von
Approved by

Lucas HOŁOWICZ
Lucas Holowicz

20. JAN. 2022

Datum / Date: