

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2024-03
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0321258	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.43	PASS
Intercept	---	0.12	--	0.11	PASS
r	0.990	---	--	0.993	PASS
LOW RANGE					
Slope	2.8	6.9	--	3.9	PASS
Intercept	---	0.21	--	0.13	PASS
r	0.990	---	--	0.999	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0321257	3 mL
COAMATIC FVIII S-2765 + I-2581	N0321010	6 mL
COAMATIC FVIII BUFFER	N0320810	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Nehal Patel *N. M. Patel* DATE: 6/20/2022
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 6/20/2022
 PRINT SIGNATURE

TITLE: Senior Group Leader