

CERTIFICATE OF ANALYSIS




**Instrumentation
Laboratory**
 A Werfen Company

PRODUCT NAME:	COATEST APC Resistance V	EXPIRATION DATE:	2023-06
PRODUCT NUMBER:	82312063		
LOT NUMBER:	N0421451		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	Pass/Fail
	MIN	MAX			
APT-time on ACL					
Diluted Control Plasma Level 1	24	42	seconds	31	PASS
Diluted Control Plasma Level 2	25	45	seconds	31	PASS

APC-V ratio on ACL	MIN	MAX	UNITS	RESULTS	Pass/Fail
Control Plasma Level 1	2.8	3.2	--	2.9	PASS
Control Plasma Level 2	1.4	1.9	--	1.9	PASS

KIT COMPONENTS:

NAME	LOT NUMBER	VOLUME
COATEST APC Resistance CaCl ₂	N1118204	8 mL
COATEST APC Resistance APTT reagent	N1218570	16 mL
COATEST APC Resistance APC/CaCl ₂	N0128927	2 mL
COATEST APC Resistance Control Plasma Level 1	N0129155	1 mL
COATEST APC Resistance Control Plasma Level 2	N0320802	1 mL
COATEST APC Resistance V-DEF Plasma	N1218392	4 mL

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Girle Reinoso G. Reinoso DATE: 5/4/2022 5/4/2022
 PRINT SIGNATURE

TITLE: QC Analyst ???

APPROVAL SIGNATURE: D.Tencza D. Tencza DATE: 5/4/2022 5/4/2022
 PRINT SIGNATURE

TITLE: Senior Group Leader