

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME: <u>S-2222, 25 mg Substrate</u>	EXPIRATION DATE: <u>2025-02</u>
PRODUCT NUMBER: <u>82031639</u>	
LOT: <u>N1017109</u>	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY:

Maria Somoza
PRINT

Maria Somoza
SIGNATURE

DATE: 12/20/2021

12/20/21

TITLE: Quality Control II

APPROVAL SIGNATURE:

Donna Tencza
PRINT

Donna Tencza
SIGNATURE

DATE: 12/20/2021

12/20/21

TITLE: Senior Group Leader

QCM00820316-2 Rev. 00