



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ2K00.02

**FVIII Inhibitorplasma HCV neg. 5x1mL**

REF 5159010-RUO

LOT 2K91B02.03

 2023-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.4	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	FVIII clotting Siron LS	14.9 – 24.9 BE/mL	24.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
Lucas HOLOVICZ

25. JULI 2022

Datum / Date:

Freigabe von  
Approved by

  
Michael KAEKA

25. JULI 2022

Datum / Date: