

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	<u>FXa, 10 x 71 nkat</u>	EXPIRATION DATE:	<u>2023-09</u>
PRODUCT NUMBER:	<u>82098539</u>		
LOT NUMBER:	<u>N0311444</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					PASS/FAIL
ACTIVITY	61	81	nkat/vial	72	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	0.999	PASS
CONTROL RECOVERY	65.7	72.7	nkat/vial	70.8	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	VOLUME
FXa, 71 nkat	N0311444	2 ml

PREPARED BY: MARIA OETTINGER *Maria Oettinger*
PRINT SIGNATURE

DATE: 4-6-2021

TITLE: QC ANALYST III

APPROVAL SIGNATURE: DEBRA HARZ *Debra Harz*
PRINT SIGNATURE

DATE: 4-6-2021

TITLE: QC TECHNICAL SUPERVISOR